

History

(Total System Time: **6d 0h 17m**)


Event	System start	Cartridge installation (A,B)	Cells injection	Cartridge installation (C)	Cartridge installation (D)	Harvesting Bottle Disconnection
Date	2018-03-28, 13:55	2018-03-28, 13:59	2018-03-28, 15:05	2018-04-03, 13:50	---	2018-04-03, 14:13
User ID	VBC-002-MMOR	VBC-002-MMOR	VBC-005-FVUL	VBC-002-MMOR	---	VBC-002-MMOR
Code		001201709252A 001201709372B		001201709442C	---	

In-Process Parameters

Flask Temperature: **36,79°C ±0,10°C**

Cell Culture Medium pH: **7,19 ±0,04**

Cell Confluency



Samples

Events

Description	Phase	Date	User ID
Open door	9.3	2018-03-04, 13:43	VBC-002-MMOR
Skip forward	9.3	2018-03-04, 13:49	VBC-002-MMOR
Skip forward	9.4	2018-03-04, 13:49	VBC-002-MMOR
Skip forward	9.5	2018-03-04, 13:49	VBC-002-MMOR

Warnings

Description	Phase	Date	Duration
Gas flow error	3.3	2018-03-29, 14:29	0h 2m
Lid open overtime	3.3	2018-03-29, 14:34	0h 8m
Unable to detect CCM	5.3	2018-03-29, 16:03	0h 1m
Lid open overtime	9.3	2018-04-03, 13:39	0h 3m